



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

on

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/694,068	10/27/2003	Donald W. Howie	0103-0035(ZM0476)	3820

43231 7590 09/29/2004

ZIMMER TECHNOLOGY - REEVES  
P. O. BOX 1268  
ALED0, TX 76008

EXAMINER

SNOW, BRUCE EDWARD

ART UNIT PAPER NUMBER

3738

DATE MAILED: 09/29/2004

Please find below and/or attached an Office communication concerning this application or proceeding.

<b>Office Action Summary</b>	<b>Application No.</b>	<b>Applicant(s)</b>	
	10/694,068	HOWIE ET AL.	
	<b>Examiner</b>	<b>Art Unit</b>	
	Bruce E Snow	3738	

**-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --**

**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☐ Responsive to communication(s) filed on \_\_\_\_.
- 2a) ☐ This action is **FINAL**.                      2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1-24 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-24 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
     Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
     Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- |  |   |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)                        | 4) <input type="checkbox"/> Interview Summary (PTO-413)                     |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)               | Paper No(s)/Mail Date. ____.  |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| Paper No(s)/Mail Date <u>10272003</u> .  | 6) <input type="checkbox"/> Other: ____.                                    |

## **DETAILED ACTION**

### ***Claim Rejections - 35 USC § 112***

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claim 6 is rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

Regarding claim 6, member lacks antecedent basis.

### ***Specification***

The specification is objected to as failing to provide proper antecedent basis for the claimed subject matter. See 37 CFR 1.75(d)(1) and MPEP § 608.01(o). Correction of the following is required: Most of applicant's claim terminology is not supported in the specification making claim interpretation impossible. For example, means for preventing, means for anchoring, member, means for spacing, engagement member, body, retention member, anchor portion are not supported.

As far as the claims can be determined, the following is made of record.

### ***Claim Rejections - 35 USC § 102***

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States

Art Unit: 3738

only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

(e) the invention was described in a patent granted on an application for patent by another filed in the United States before the invention thereof by the applicant for patent, or on an international application by another who has fulfilled the requirements of paragraphs (1), (2), and (4) of section 371(c) of this title before the invention thereof by the applicant for patent.

The changes made to 35 U.S.C. 102(e) by the American Inventors Protection Act of 1999 (AIPA) and the Intellectual Property and High Technology Technical Amendments Act of 2002 do not apply when the reference is a U.S. patent resulting directly or indirectly from an international application filed before November 29, 2000. Therefore, the prior art date of the reference is determined under 35 U.S.C. 102(e) prior to the amendment by the AIPA (pre-AIPA 35 U.S.C. 102(e)).

Claims 1-4, 6, 15-17, 24 are rejected under 35 U.S.C. 102(b) as being anticipated by Carpenter et al (5,425,768).

Carpenter et al teaches a positioner for implantation adjacent a femoral hip implant for retaining the femoral hip implant in a femoral canal having a canal wall, the positioner comprising:

means for preventing the femoral hip implant from rising out of the femoral canal beyond a predetermined position while permitting subsidence of the femoral hip implant down into the femoral canal Including elements 46 which couples to the stem 22 of the femoral hip implant; and

means for anchoring the means for preventing relative to the femoral canal 62 which are capable of being wedged in the femoral canal providing anchoring and further provide a "means for anchoring" once the cement cures.

Regarding the means for spacing, see element 48.

Claims 1-10 and 15-24 are rejected under 35 U.S.C. 102(b) as being anticipated by Link et al (4,698,063).

Link et al teaches a positioner for implantation adjacent a femoral hip implant for retaining the femoral hip implant in a femoral canal having a canal wall, the positioner comprising:

means for preventing the femoral hip implant from rising out of the femoral canal beyond a predetermined position while permitting subsidence of the femoral hip implant down into the femoral canal Including elements 10 and serrations;

and means for anchoring the means for preventing relative to the femoral canal 13 and the serrations the stem.

Regarding claim 6, see the serrations on the hip implant.

The spacing elements/means/etc. include elements 13 and the fins on which the serrations are on. See figure 2 showing anterior and posterior fins.

Regarding claim 7, see first member 10 and second member 13.

Regarding claim 15, anchor portion, see at least the serrations.

Claim 17, see at least the fins.

Claim 18, "retention member" is element 10.

Claim 22, elements 13 and 10 form a symmetrical "L" shape.

Claims 1-7, 15-20, 22-23 are rejected under 35 U.S.C. 102(b) as being anticipated by Mathys, Sr. et al.

Art Unit: 3738

Mathys, Sr. et al teaches a positioner for implantation adjacent a femoral hip implant for retaining the femoral hip implant in a femoral canal having a canal wall, the positioner comprising:

means for preventing the femoral hip implant from rising out of the femoral canal beyond a predetermined position while permitting subsidence of the femoral hip implant down into the femoral canal (see screw head);

and means for anchoring the means for preventing relative to the femoral canal (see screw 9).


Any inquiry concerning this communication or earlier communications from the examiner should be directed to Bruce E Snow whose telephone number is (703) 308-3255. The examiner can normally be reached on Mon-Thurs.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott can be reached on (703)308-2111. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Art Unit: 3738

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

bes



BRUCE SNOW  
PRIMARY EXAMINER